



199 Fire Tower Drive  
Tonawanda, NY USA

Phone: (716) 629-3800  
FAX: (716) 693-9162



## Certificate of Compliance

The following models, with the MD option:

**544, and 574**

Were evaluated by:

FM Approvals  
1151 Boston-Providence Turnpike  
P.O. Box 9102  
Norwood, MA 02062, USA

They were found to comply with the following United States Regulations:

Approval Standard Class No 3600: 1998  
Approval Standard Class No. 3610:2010  
Approval Standard Class No. 3810: 2005  
NEMA 250: 2008

This shows that these products are Intrinsically Safe and Entity & Suitable for use in the following Hazardous Locations:

**Class I, II, and III, Division 1, Groups A, B, C, D, E, F, and G**  
NEMA 4X (indoor and outdoor)

When installed per the following drawings:

544 and 574 [CD0602](#)

This is specified on the FM Certificate of Compliance [3Y5A1.AX](#)

The products are marked with the FM Mark and the above ratings

I declare that the above statements are factual.

10/16/17

---

Joel Neri  
Viatran Quality/Approvals Manager

---

Date

98MD5X4COFC Rev. C