



199 Fire Tower Drive  
Tonawanda, NY USA

Phone: (716) 629-3800  
FAX: (716) 693-9162



## Certificate of Compliance

The following models, with the ME option:

**570, 571, 574, 870, and 970**

Were evaluated by:

CSA International  
8501 E. Pleasant Valley Road  
Cleveland, OH 44131

They were found to comply with the following Canadian Regulations:

CAN/CSA Standard C22.2 No. 0: M91 (Reaffirmed 2001)  
CSA Standard C22.2 No. 25: 1966  
CSA Standard C22.2 No. 30: 1986  
CAN/CSA Standard C22.2 No. 94: M91  
CSA Standard C22.2 No. 142: M1987  
CSA Standard C22.2 No. 1010-1: 1992  
CSA Standard C22.2 No. 1010.1B: 97  
TIL No. I-29B  
CAN/CSA—E79-0: 95  
CAN/CSA—E79-11: 95

This shows that these products are Explosion Proof and can be used in the following Hazardous Locations:

**Class I, Division 1, Groups A, B, C, and D**

**Class II, Division 1, Groups E, F, G**

**Class III**

**Enclosure Type 4, Dual Seal**

This is specified on the CSA Certificate of Compliance [1000784 \(LR 34272\)](#)

The products are marked with the CSA Mark and the above ratings

I declare that the above statements are factual.

4/9/18

---

Joel Neri  
Viatran Quality/Approvals Manager

---

Date

98MECOFC Rev. C