



199 Fire Tower Drive  
Tonawanda, NY USA

Phone: (716) 629-3800  
FAX: (716) 693-9162



## EU Declaration of Conformity

The following models, with the NG option:

**548, 570, 571, 572, 584, 748, 770, 848, 870, 948, 970, LP 748, and LP 770**

Comply with the following EU Directives

EMC Directive 2014/30/EC

EN 61326-1:2013

EN 55011:2009+A1:2010

EN 61000-4-2:2009

EN 61000-4-3:2006+A2:2010

EN 61000-4-4:2012

EN 61000-4-5:2014

EN 61000-4-6:2014

ATEX Directive 2014/34/EU

EN 60079-0: 2012 + A11: 2013

EN 60079-1: 2014

EC-Type Examination Certificate Number: [Presafe 16 ATEX 8250X](#)

This shows that these products are flameproof and can be used in the following Hazardous Locations:

**II 2G Ex db IIC T6...T4 Gb -20°C ≤ Ta ≤ +60°C**

Notified Body            DNV Nemko Presafe AS  
2460                        Veritasveien 3  
1363 Høvik, Norway

The products are marked with the CE Mark and the above ratings

I declare the above products meet the essential requirements that are specified above.

4/9/21

Joel Neri  
Viatran Quality/Approvals Manager

\_\_\_\_\_  
Date