

199 Fire Tower Drive Tonawanda, NY USA

Phone: (716) 629-3800 FAX: (716) 693-9162



EU Declaration of Conformity

The following models, with the NK option: 510 and 520

Comply with the following EU Directives EMC Directive 2014/30/EU EN 61326-1:2013 EN 55011:2009+A1:2010 EN 61000-4-2:2009 EN 61000-4-2:2009 EN 61000-4-3:2006+A2:2010 EN 61000-4-4:2012 EN 61000-4-5:2014 EN 61000-4-6:2014 ATEX Directive 2014/34/EU EN 60079-0: 2012 +A11:2013 EN 60079-11: 2012

EU-Type Examination Certificate Number: Presafe 16 ATEX 8465X

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

E II 1 G Ex ia IIC Ga Temperature Code T4: -20°C \leq Ta \leq +80°C Temperature Code T5: -20°C \leq Ta \leq +40°C

When installed per drawing: CD0666

Notified Body 2460 Det Norske Veritas Certification AS Veritasveien 3 1363 Høvik, Norway

PED 2014/68/EU

Viatran Corporation's evaluation of conformity to the PED Directive, Category 1, based on a Module A assessment. This applies to all transmitters listed above for Maximum Working Pressure ranges equal to or greater than 200 bar (2900 PSI).

RoHS EU 2015/863

This product is RoHS compliant.

The products are marked with the CE Mark and the above ratings

I declare the above products meet the essential requirements that are specified above.

Ohen

Joel Neri Viatran Quality/Approvals Manager <u>4/9/2021</u> Date