



199 Fire Tower Drive  
Tonawanda, NY USA

Phone: (716) 629-3800  
FAX: (716) 693-9162



## EU Declaration of Conformity

The following models, with the NK option:

**521, 522**

Comply with the following EU Directives

EMC Directive 2014/30/EU

EN 61326-1:2013

EN 55011:2009+A1:2010

EN 61000-4-2:2009

EN 61000-4-3:2006+A2:2010

EN 61000-4-4:2012

EN 61000-4-5:2014

EN 61000-4-6:2014

ATEX Directive 2014/34/EU

EN 60079-0: 2012 + A11: 2013

EN 60079-11: 2012

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

 **II 1 G Ex ia IIC Ga**

Temperature Code T4:  $-40^{\circ}\text{C} \leq T_a \leq +85^{\circ}\text{C}$

Temperature Code T5:  $-40^{\circ}\text{C} \leq T_a \leq +40^{\circ}\text{C}$

When installed per drawing: [CD0676](#)

Notified Body      FM Approvals Europe Ltd.  
2809                    1 Georges Quay Plaza  
Dublin Ireland D02 E440

This is specified in detail on:

EU-Type Examination Certificate Number: [FM16ATEX0043](#)

PED 2014/68/EU

Viatran Corporation's evaluation of conformity to the PED Directive, Category 1, based on a Module A assessment. This applies to all transmitters listed above for Maximum Working Pressure ranges equal to or greater than 200 bar (2900 PSI).

RoHS EU 2015/863

This product is RoHS compliant.

The products are marked with the CE Mark and the above ratings

I declare the above products meet the essential requirements that are specified above.

Joel Neri  
Viatran Quality Manager

4/9/21

Date