



199 Fire Tower Drive
Tonawanda, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



EU Declaration of Conformity

The following models, with the NK option:

570, 571 and 572

Comply with the following EU Directives

EMC Directive 2014/30/EU

EN 61326-1:2013

EN 55011:2009+A1:2010

EN 61000-4-2:2009

EN 61000-4-3:2006+A2:2010

EN 61000-4-4:2012

EN 61000-4-5:2014

EN 61000-4-6:2014

ATEX Directive 2014/34/EU

EN 60079-0: 2012; A11: 2013

EN 60079-11: **2012**

EC-Type Examination Certificate Number: [LCIE 03 ATEX 6373 X-03](#)

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

 **II 1 G Ex ia IIC Ga**

T4: $-20^{\circ}\text{C} \leq \text{Ta} \leq 75^{\circ}\text{C}$

T5: $-20^{\circ}\text{C} \leq \text{Ta} \leq 40^{\circ}\text{C}$

And for the 571 only

T3 $-20^{\circ}\text{C} \leq \text{Ta} \leq 88^{\circ}\text{C}$

When installed per drawings:

[CD0628](#) 570, 571 and 572

Notified Body LCIE Bureau Veritas
33 av. du Général Leclerc
92260, Fontenay aux Roses - France

The products are marked with the CE Mark and the above ratings

I declare the above products meet the essential requirements that are specified above.

11/8/17

Joel Neri
Viatran Quality/Approvals Manager

Date