



199 Fire Tower Drive
Tonawanda, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



EU Declaration of Conformity

The following models, with the NK option:

548, 748, 770, 848, 870, LP 748, and LP 770

Comply with the following EU Directives

EMC Directive 2014/30/EU

EN 61326-1:2013

EN 55011:2009+A1:2010

EN 61000-4-2:2009

EN 61000-4-3:2006+A2:2010

EN 61000-4-4:2012

EN 61000-4-5:2014

EN 61000-4-6:2014

ATEX Directive 2014/34/EU

EN 60079-0: 2012 +A11 2013

EN 60079-11:2012

EU Type Examination Certificate Number: [Presafe 16 ATEX 8249X](#)

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

II 1 G Ex ia IIC Ga

Temperature Code T4 -20 C ≤ Ta ≤ 80 C

When installed per drawings:

[CD0645](#) 548, 848, 870

[CD0651](#) 748, 770, LP748, LP 770

Notified Body DNV Nemko Presafe AS

2460 Veritasveien 3

1363 Høvik, Norway

The products are marked with the CE Mark and the above ratings

I declare the above products meet the essential requirements that are specified above.

4/9/21

Joel Neri
Viatran Quality/Approvals Manager

Date