

199 Fire Tower Dr. Tonawanda NY 14150 USA Phone: (716) 629-3800 FAX: (716) 693-9162



Certificate of Compliance

The following models, with the NX option: 548, 748, 770, LP748 and LP770

Were evaluated by: CSA International 8501 E. Pleasant Valley Road Cleveland, OH 44131

They were found to comply with the following Canadian Regulations:

CAN/CSA Standard C22.2 No. 0- M91 CSA Std C22.2 No. 142-M1987 CAN/CSA-C22.2 No. 157- 92 CSA Std C22.2 No.213-M1987 CSA Std C22.2 No. 94- M91 CAN/CSA – E60079-0-02 CAN/CSA – E60079-11-02

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

Class I, Division 1, Groups A, B, C, and D Class II, Division 1, Groups E, F, G (except Pressure Format G) Class III, Division 1 (except Pressure Format G) Ex ia IIC; Class I, Zone 0 Enclosure Type 4, Dual Seal (except Pressure Format G) Temperature Code T4 at Tamb = 80°C Temperature Code T5 at Tamb = 40°C

When installed per the following drawings:

CD0646548CD0649748, 770, LP748, LP770

This is specified on the CSA Certificate of Compliance 1507558 (LR 34272)

The products are marked with the CSA Mark and the above ratings

I declare that the above statements are factual.

4/27/2021 Date

Quality/Approvals Manager

Page 1 of 1