



3829 Forest Parkway, Suite 500
Wheatfield, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Compliance

The following models, with the NZ option:

948 and 970

Were evaluated by:

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102
Norwood, MA 02062, USA

They were found to comply with the following United States Regulations:

Approval Standard Class No 3600: 1998
Approval Standard Class No. 3611: 2004
Approval Standard Class No. 3810: 1989; Supplement 1: 1995
NEMA 250: 1991

This shows that these products are nonincendive and can be used in the following Hazardous Locations:

Class I, Division 2, Groups A, B, C, and D
Class II, and III, Division 2; Groups F and G
Class I, Zone 2 Group IIC
Dual Seal, NEMA Type 4X
Temperature Code T4 in and ambient of 60°C or less

This is specified on the FM Certificate of Compliance [3009321](#)

The products are marked with the FM Mark and the above ratings

I declare that the above statements are factual.

Bruce R. Temple
Certification/Regulation Lead