



3829 Forest Parkway, Suite 500
Wheatfield, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Compliance

The following models, with the TF option:
509, 509T20B, 509T20E, and 709

Were evaluated by:
FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102
Norwood, MA 02062, USA

They were found to comply with the following United States Regulations:
Approval Standard Class No 3600: 1998
Approval Standard Class No. 3610:1999
Approval Standard Class No. 3810: 1989

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:
Class I, Division 1, Groups A, B, C, D
Class I, Zone 0 Group IIC AEx ia
Temperature Code T5 in an ambient of 40°C or less
NEMA 4X (indoor and outdoor)

When installed per the following drawings:
[CD0641](#)

This is specified on the FM Certificate of Compliance [3018059](#)

The products are marked with the FM Mark and the above ratings

I declare that the above statements are factual.

Bruce R. Temple
Certification/Regulation Lead