



3829 Forest Parkway, Suite 500
Wheatfield, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Compliance

The following models, with the TF option:

948, 970

Were evaluated by:

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102
Norwood, MA 02062, USA

They were found to comply with the following United States Regulations:

Approval Standard Class No 3600: 1998
Approval Standard Class No. 3610:1999
Approval Standard Class No. 3810: 1989
Supplement #1: 1995
NEMA 250: 1991

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

Class I, II, and III, Division 1, Groups A, B, C, D, E, F, G
Class I, Zone 0, AEx ia IIC
Temperature Code T4 in an ambient of 60°C or less
NEMA 4X (indoor and outdoor)
Dual Seal

When installed per the following drawings:

[CD0636](#)

This is specified on the FM Certificate of Compliance [3009321](#)

The products are marked with the FM Mark and the above ratings

I declare that the above statements are factual.

Bruce R. Temple
Certification/Regulation Lead