



3829 Forest Parkway, Suite 500
Wheatfield, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Compliance

The following models, with the TF option:

350, 351, 550, 551, 553, 554, 555, 570, 571, 572, and 584

Were evaluated by:

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102
Norwood, MA 02062, USA

They were found to comply with the following United States Regulations:

Approval Standard Class No 3600: 2011
Approval Standard Class No. 3610:2010
Approval Standard Class No. 3810: 2005
ANSI/ISA-60079-0::2009
ANSI/ISA-60079-11:2009

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

Class I, II, and III, Division 1, Groups A, B, C, D, E, F, and G

Class I, Zone 0 Group IIC

Temperature Code T4 in and ambient of 80°C or less
Temperature Code T5 in an ambient of 40°C or less
NEMA 4X (indoor and outdoor)

When installed per the following drawings:

350, and 351,	<u>CD0602</u>
550, 551, 553, 554, and 555	<u>CD0667</u>
570, 571, 572, and 584	<u>CD0633</u>

This is specified on the FM Certificate of Compliance


350, 351	<u>3035562</u>
550, 551, 553, 554, 555, 570, 571, 572, and 584	<u>3V0A8.AX</u>

The products are marked with the FM Mark and the above ratings

I declare that the above statements are factual.



Joel Neri
Engineering Manager



Date