



199 Fire Tower Drive
Tonawanda, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Compliance

The following models, with the TG option:

521, 522

Were evaluated by:

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102
Norwood, MA 02062, USA

They were found to comply with the following standards:

CAN/CSA- C22.2 No. 213- M1987:2013	FM Class 3610: 2015
CAN/CSA- C22.2 No. 157- 92: 2012	FM Class 3810: 2005
CAN/CSA- C22.2 1010.1: 2004	ANSI/ISA 61010-1: 2004
CAN/CSA- C22.2 No. 60079-0: 2015	ANSI/ISA 60079-0: 2013
CAN/CSA- C22.2 No. 60079-11: 2014	ANSI/ISA 60079-11: 2014
FM Class 3600: 2018	

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

Class I, Division 1, Groups A, B, C, and D

Class I, Zone 0 Ex ia IIC

Temperature Code T4 Ta = -40°C to 85°C

Temperature Code T5 Ta = -40°C to 40°C

When installed per the following drawings: [CD0676](#)

This is specified in detail on:

FM Certificate of Conformity: [FM16US0170](#)

FM Certificate of Conformity: [FM16CA0100](#)

The products are marked with the FM Mark and the above ratings

I declare that the above statements are factual.

4/20/21

Joel Neri
Viatran Quality/Approvals Manager

Date