



199 Fire Tower Dr  
Tonawanda, NY 14150 USA

Phone: (716) 629-3800  
FAX: (716) 693-9162



## Certificate of Compliance

The following models, with the TG option:  
**517, 817, and WW517**

Were evaluated by:

CSA International  
8501 E. Pleasant Valley Road  
Cleveland, OH 44131

They were found to comply with the following Canadian Regulations:

CSA Standard C22.2 No. 142: M1987  
CAN/CSA Standard C22.2 No. 157: 92  
CAN/CSA E79-0: 95  
CAN/CSA E79-11: 95

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

**Class I, Division 1, Groups A, B, C, and D**  
**Class I, Zone 0 Ex ia IIC**  
Temperature Code T4 at Ta=80°C

When installed per the following drawings:

[CD0635](#) 517, 817 and WW517

This is specified on the CSA Certificate of Compliance [1300108](#)

The products are marked with the CSA Mark and the above ratings

I declare that the above statements are factual.

Joel Neri  
Quality/Approvals Manager

4/26/2021

Date