



199 Fire Tower Drive
Tonawanda, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Compliance

The following models, with the TJ option:

511, 512, and 513

Were evaluated by:

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102
Norwood, MA 02062, USA

They were found to comply with the following standards:

CSA- C22.2 No. 213-M1987: R2013
CSA- C22.2 No. 157-92: R2012
CSA- C22.2 1010-1: 2004
CAN/CSA- C22.2 No. 60079-0: 2011
CAN/CSA- C22.2 No. 60079-11: 2014
CAN/CSA- C22.2 No. 60079-15: 2012
CSA- C22.2 No. 60529: R2010
CSA- C22.2 No. 94-M91: R2011

This shows that these products are Nonincendive and can be used in the following Hazardous Locations:

Class I, Division 2, Groups A, B, C, and D

Class I, Zone 1, Ex nA IIC

Temperature Code T4 Ta = -40°C to 85°C

Temperature Code T5 Ta = -40°C to 40°C

When installed per the following drawings:

[CD0673](#)

Enclosure rating of NEMA Type 6P and IP68

This is specified on the FM Certificate of Compliance [FM16CA0133X](#)

The products are marked with the FMc Mark and the above ratings

I declare that the above statements are factual.

4/20/2021

Joel Neri

Date

Viatran Quality/Approvals Manager