

199 Fire Tower Drive Tonawanda, NY USA

Phone: (716) 629-3800 FAX: (716) 693-9162



## **EU Declaration of Conformity**

The following models, with the TK option: 548, 570, 571, 748, 770, 848, 870, LP 748, and LP 770

Comply with the following EU Directives

EMC Directive 2004/108/EC

EN 61326-1:2013

EN 55011:2009+A1:2010

EN 61000-4-2:2009

EN 61000-4-3:2006+A2:2010

EN 61000-4-4:2012

EN 61000-4-5:2014

EN 61000-4-6:2014

ATEX Directive 2014/34/EU

EN 60079-0:2012 + A11 2013

EN 60079-15:2010

EC-Type Examination Certificate Number: Presafe 16 ATEX 8251X

This shows that these products are non-sparking and can be used in the following Hazardous Locations:

**ⓑ** II 3 G Ex nA IIC T4 Gc -20°C ≤  $T_a$  ≤ 80°C

Notified Body DNV Nemko Presafe AS

Gaustadalléen 30 0373 Oslo, Norway

The products are marked with the CE Mark and the above ratings

I declare the above products meet the essential requirements that are specified above.

Joel Neri

Joel Neri

Date

Viatran Quality/Approvals Manager