



199 Fire Tower Drive
Tonawanda, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Conformity

The following models, with the TW option (includes FA option):

509, 709, 510, 520, 521, 522

Technical Regulations

TP TC 012/2011

GOST 31610.0-2014 (IEC 60079-0: 2011)

GOST 31610.11-2014 (IEC 60079-11:2014)

GOST 31610.15-2014 (IEC 60079-15:2010)

GOST IEC 60079-1-2011

GOST IEC 60079-10-1-2011

GOST IEC 60079-14-2013

GOST R IEC 60079-20-1-2011

GOST 14254-2015 (IEC 60259:2013)

Certificate Number: [TC RU C-US.BH02.B.00653/18](#)

This shows that these products are Intrinsically Safe and can be used in the following
Hazardous Locations:

0Ex ia IIC Ga

Temperature Code T4: $-20^{\circ}\text{C} \leq T_a \leq +80^{\circ}\text{C}$

Temperature Code T5: $-20^{\circ}\text{C} \leq T_a \leq +40^{\circ}\text{C}$

Metrology Certificate: [OC.C.30.001.A No 73774](#)

Registration Number: 74921-19

Valid Through: April 29th, 2024

Notified Body

VNIIFTRI

Federal Agency on Technical Regulation and Metrology

141570 Mendeleevo

Moscow Region, Russia

I declare the above products meet the essential requirements that are specified above.


Joel Neri

Viatran Quality/Approvals Manager

7/24/2019

Date