



199 Fire Tower Drive
Tonawanda, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Conformity

The following models, with the TW option:

570, 770, 870, 571, 548, 748, 848, LP748

Technical Regulations

TP TC 012/2011
GOST 31610.0-2014 (IEC 60079-0: 2011)
GOST 31610.11-2014 (IEC 60079-11:2014)
GOST 31610.15-2014 (IEC 60079-15:2010)
GOST IEC 60079-1-2011
GOST IEC 60079-10-1-2011
GOST IEC 60079-14-2013
GOST R IEC 60079-20-1-2011
GOST 14254-2015 (IEC 60259:2013)

Certificate Number: [TC RU C-US.BH02.B.00653/18](#)

This shows that these products are Intrinsically Safe, Flameproof, and Non-Sparking and can be used in the following Hazardous Locations:

0Ex ia IIC Ga X – (all products)
Temperature Code T4: $-20^{\circ}\text{C} \leq T_a \leq +80^{\circ}\text{C}$
Temperature Code T5: $-20^{\circ}\text{C} \leq T_a \leq +40^{\circ}\text{C}$
1Ex d IIC Gb X – (standard, ZL, ZU only)
Temperature Code T6: $-20^{\circ}\text{C} \leq T_a \leq +40^{\circ}\text{C}$
2Ex nA IIC Gc X – (all products)
Temperature Code T4: $-20^{\circ}\text{C} \leq T_a \leq +80^{\circ}\text{C}$

Notified Body **VNIIFTRI**
Federal Agency on Technical Regulation and Metrology
141570 Mendeleevo
Moscow Region, Russia

I declare the above products meet the essential requirements that are specified above.



Joel Neri
Viatran Quality/Approvals Manager

2/28/2019

Date