



199 Fire Tower Drive
Tonawanda, NY 10150

Phone: (716) 629-3800
FAX: (716) 693-9162



EU Declaration of Conformity

The part number 5715YYSX1670, where YY stands for two letters that indicate the sensor range

Comply with the following EU Directives

EMC Directive 2014/30/EU

EN 61326-1:2006

EN 55011:2009+A1:2010

EN 61000-4-2:2009

EN 61000-4-3:2006+A2:2010

EN 61000-4-4:2004+A1:2010

EN 61000-4-5:2006

EN 61000-4-6:2009

PED 2014/68/EU

PED Directive, Category III, based on a Modules B1 and D assessment.

Certificate Number:

[4037-214-CE-USA-DNV](#) and
[9632-2017-CE-USA-ACCREDIA](#)

The wetted parts are made with

[HOU-PMA-13-201 \(PMA Ref. No. 005711\)](#) or
[HOU-PMA-13-202 \(PMA Ref. No. 020559\)](#)

Notified Body

DNV GL Business Assurance Italia S.r.l.
Via Energy Park, 14
20871 Vimercate (MB), Italy

ATEX Directive 2014/34/EU

EN 60079-0: 2009

EN 60079-11: 2012

EC-Type Examination Certificate Number: [LCIE 03 ATEX 6373 X](#)

This shows that these products are Intrinsically Safe and can be used in the following
Hazardous Locations:

 II 1 G Ex ia IIC Ga

T3: $-20^{\circ}\text{C} \leq T_a \leq 88^{\circ}\text{C}$

T4: $-20^{\circ}\text{C} \leq T_a \leq 75^{\circ}\text{C}$

T5: $-20^{\circ}\text{C} \leq T_a \leq 40^{\circ}\text{C}$

When installed per drawings:

[CD0628](#)

571

Notified Body

LCIE Bureau Veritas
33 av. du Général Leclerc
92260, Fontenay aux Roses - France

I declare the above products, when marked with the CE Mark; meet the essential requirements that are specified above.



Joel Neri

Viatran Quality / Approvals Manager

4/2/2018

Date